



December 19, 2018

**S U P P L E M E N T A L / B I D B U L L E T I N**  
**No. 10 – 2018**

**19PBGNMSC02.2 - G**  
**Supply & Delivery of BRRIM Equipment – AQUA MARINE**  
PhilGEPS Reference No.: 5896897

This Supplemental Bid Bulletin is issued to modify or amend the item in the Bid Documents of the above-captioned project.

Reference	AMENDMENT/CHANGE/ CLARIFICATION
<b>VII. Technical Specifications</b> <b>Item No. 9</b> 1 unit Refractometer	5 units Refractometer Specifications: Hand Held Refractometer Dual Scale Salinity Measure range: 0-100%(0-10%) Accuracy: 0.001sg (1%) Salinity Proportion: 1.000-1.070sg) Features: Used to measure the salinity and specific gravity of the water High Quality and Accurate testing result Easily calibrated Automatic temperature compensation
<b>Item No. 10</b> Dissolve Oxygen Meter	<b>5 units Dissolved Oxygen Meter</b> Specifications: Pen Type Oxygen Meter Range: 0-20 mg/L Accuracy: 0.1 mg/L Basic Error: ±0.3 mg/L Dot matrix backlit LCD display Automatic temperature compensation, fast reponse time With the dissolved oxygen electrode (Polarographic film type)

This Supplemental/Bid Bulletin shall form part of the documents under this procurement. Any provisions in the terms of reference inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

  
**AMYTHYEST M. TALABA, MPA**  
BAC Chairman